



Patent and Trademark Office  
Address: COMMISSIONER OF PATENTS AND TRADEMARKS  
Washington, D.C. 20231

APPLICATION NUMBER	FILING DATE	FIRST NAMED APPLICANT	ATTY. DOCKET NO.
09/152,195	09/14/98	CLARK	J

EXAMINER

J DAVID NELSON  
10885 SOUTH STATE STREET  
SANDY UT 84070

QM12/1028

LACYK, J	PAPER NUMBER
ART UNIT	

3736

DATE MAILED: 10/28/99

This is a communication from the examiner in charge of your application.  
COMMISSIONER OF PATENTS AND TRADEMARKS

### OFFICE ACTION SUMMARY

- ☐ Responsive to communication(s) filed on \_\_\_\_\_
- ☐ This action is FINAL.
- ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 D.C. 11; 453 O.G. 213.

A shortened statutory period for response to this action is set to expire 3 month(s), or thirty days, whichever is longer, from the mailing date of this communication. Failure to respond within the period for response will cause the application to become abandoned. (35 U.S.C. § 133). Extensions of time may be obtained under the provisions of 37 CFR 1.136(a).

#### Disposition of Claims

- ☒ Claim(s) 1-9 is/are pending in the application.  
Of the above, claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- ☒ Claim(s) 1-9 is/are rejected.
- ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- ☐ Claim(s) \_\_\_\_\_ are subject to restriction or election requirement.

#### Application Papers

- ☐ See the attached Notice of Draftsperson's Patent Drawing Review, PTO-948.
- ☐ The drawing(s) filed on \_\_\_\_\_ is/are objected to by the Examiner.
- ☐ The proposed drawing correction, filed on \_\_\_\_\_ is ☐ approved ☐ disapproved.
- ☐ The specification is objected to by the Examiner.
- ☐ The oath or declaration is objected to by the Examiner.

#### Priority under 35 U.S.C. § 119

- ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d).
- ☐ All ☐ Some\* ☐ None of the CERTIFIED copies of the priority documents have been
- ☐ received.
- ☐ received in Application No. (Series Code/Serial Number) \_\_\_\_\_
- ☐ received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\*Certified copies not received: \_\_\_\_\_

- ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e).

#### Attachment(s)

- ☐ Notice of Reference Cited, PTO-892
- ☒ Information Disclosure Statement(s), PTO-1449, Paper No(s). 2
- ☐ Interview Summary, PTO-413
- ☐ Notice of Draftsperson's Patent Drawing Review, PTO-948
- ☐ Notice of Informal Patent Application, PTO-152

--SEE OFFICE ACTION ON THE FOLLOWING PAGES--

Art Unit: 3736

1. The articles listed on the information disclosure statement have not been considered at this time since there were no copies submitted and copies are not readily available to the examiner.

2. The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

3. Claims 1-91 are rejected under 35 U.S.C. 112, first paragraph, as containing subject matter which was not described in the specification in such a way as to enable one skilled in the art to which it pertains, or with which it is most nearly connected, to make and/or use the invention. The specification discloses that the device and method are for therapy on a body for treating sensitivities to various elements, however the specification fails to specify exactly what ailments or problems the body is being treated for or how application of the device to stimulate responses that correspond to the actual response to the product would treat the body. If the body has a sensitivity to the product, then a signal applied to the body would merely produce the same sensitivity, making it unclear what therapy is being provided.

Applicant is reminded that no new matter may be added.


4. Further applicant claims computer software for operating the device and method. It appears that a copy of the computer programs are a necessary part of the invention and should be part of the specification.

Art Unit: 3736

5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Lacyk whose telephone number is (703) 308-2995.

J.P. Lacyk

October 28, 1999

  
JOHN P. LACYK  
PRIMARY EXAMINER